

CALIBRATION ANVIL

STANDARDS: EN 12504-2, ASTM D5873, C805

Used for the verification of the calibration of the hammers.

The EN 12504:2 Specification requires obligatory the use of the anvil for the hammer tests.

The Standard specifies; before a sequence of tests on a concrete surface, take and record readings using the steel reference anvil and check to ensure that they are within the range recommended by the manufacturer. If they are not, clean and/or adjust the hammer.

After tests, take readings using the steel anvil, record them and compare them with those taken prior to the test. If the results differ, clean and/or adjust the hammer and repeat the test.

Made of hardened steel according to the standards.



HR-C7000

Technical Specifications:

Product Code	Product Name	Dimensions (cm)	Weight (kg)
HR-C7000	Calibration Anvil	Ø 15x23	16

ULTRASONIC PULSE VELOCITY AND PULSE ECHO TESTING OF CONCRETE

STANDARDS: EN 12504-4, ASTM C 597-02, BS 1881 Part 203, ISO 1920-7:2004

Structural defects cause serious damages and collapses. Ultrasonic testing provides information on the strength and uniformity of concrete, rock, composites, ceramics, wood, epoxy, refractory materials and can be used to detect and localize voids, pipes, cracks and defects.

The pulse velocity in a material depends on its density and its elastic properties which in turn are related to the quality and the compressive strength of the concrete. It is therefore possible to obtain information about the properties of components by sonic investigations.

Proceq offers the most versatile instrument for ultrasonic testing of concrete.

Ultrasonic Pulse Velocity - Pundit Lab

Measurement performance; Optimized pulse shaping, automated transmission settings for optimum performance and a range of new, more powerful transducers ensure accurate, stable measurements.

Integrated waveform display; Allows analysis of the received signal and manual triggering directly on the instrument.

On-line data acquisition; Full remote control of all transmission parameters, data logging function and functionality that turns your PC into an oscilloscope.

USB interface and data analysis software; Data analysis and export to third party programs.

Open interface; Control Pundit Lab using third party software such as LabVIEW. Pundit Lab consisting of: Display unit, 2 transducers (54kHz), 2 BNC cables 1.5 m, couplant, calibration rod, battery charger with USB-cable, 4x AA(LR6) batteries, data carrier with software, documentation and carrying case.

Ultrasonic Pulse Velocity - Pundit Lab Plus

Integrated gain stage; Removes the need for an external amplifier when using exponential transducers and long cables.



HR-C7050